

Certificate of Compliance

Shenzhen Medicalet Technology Co., Ltd
12F. Innovation Building C, Pingshan new District,
Shenzhen City 518118, China

Medicalet hereby certifies that the product identified below is inspected and released according to the requirements of product and quality specifications as maintained in our quality management system.

Product name: non-sterilized 1.5 mL Micro Centrifuge tube flat cap & non-sterilized 2.0 mL Micro Centrifuge tube flat cap
Catalog Number: MCF001015, MCF001020
Manufacture Date: 2021-10-29
Expiration Date: 2024-10-28

ATP free-This product has been tested and is free of ATP.

Non-Pyrogenic - Products labeled non-pyrogenic have been validated per FDA guidelines on LAL (Limulus Amebocyte Lysate) testing for medical devices and company guidelines.

The acceptance level for product is less than 0.5 EU/mL.

DNase & RNase Free - This product has been tested and is free of any detectable DNase / RNase contamination.

Nucleic Acid - This product has been tested and is free of any detectable Nucleic Acid.

Trace Metal - All material is certified metal-free by material manufacturers before they are turned into production which compliance with US Pharmacopoeia guidelines.

Quality Control Testing - This product has been inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release



Quality Manager

