


ESR Tube

ESR

For ESR determination without the use of sedimentation pipettes. The blood is taken directly into tubes and mixed by inversion 8-10 times. Immediately before the tubes are placed in ESR stand for measurement, the tubes must be mixed again. After one or two hours the results are read. The measurement results achieved correspond to the Westergren method.

Mixing

- ✓ 8-10 inversions
- After the blood sample has been taken is extremely important to avoid microclotting.

Classification

- ✓ Whole Blood

* According to EN, ISO standards | Quality management certified according to EN ISO 13485:2016 | CE marked

Description	Cat. No.	Cap Color	Additive	Diameter	Length	Draw Vol.
ESR PET tube	XL008130750205		Sodium Citrate 3.8%, 0.129 mol/L	13 mm	75 mm	5 mL
	XL008108120205		Sodium Citrate 3.8%, 0.129 mol/L	8 mm	120 mm	1.6 mL

* Packing info

Product	Tube Size	Primary	Secondary	Packing Size	G.W.	N.W.
ESR PET tube	13 * 75	100 pcs / tray	18 tray / Ctn	44*35*27 cm	12.7 Kg	11.5 Kg
	12 * 8	100 pcs / tray	18 tray / Ctn	44*35*27 cm	12.5 Kg	11.5 Kg